

## EU DECLARATION OF CONFORMITY

Certificate number: PS220223G

PalmSens BV declare that the product:

### EmStat Pico (V1.3 and higher):

- OEM module electrochemical analyser



conforms with the (EMC) Directive 2014/30/EU applicable standards for *Electrical equipment for measurement control and laboratory use* (standards listed below)

Standard		Class / level
EN 61326-1 (2013)	Emission	B
EN 61326-1 (2013)	Immunity	Basic EM environment

### Tests performed

Test	Test standard	Result (Pass/Fail)
Radiated emission up to 1 GHz (TEM cell)	EN 55011 (2009) + A1 (2010)	Pass
Radiated Immunity	EN-IEC 61000-4-3 (2006) + A1 (2008) + A2 (2010)	Pass
ESD	EN-IEC 61000-4-2 (2009)	Pass

Date: 23<sup>rd</sup> of February 2022



C.J. van Velzen, director

This declaration is issued under the sole responsibility of the PalmSens BV.

Instrument must be used with the recommended shielded USB cable.